

## Certificate of Testing

**Serial Number: 12511CC60A**

Page 1 of 2 Pages

FT Technologies Limited  
Church Lane  
Teddington  
Middlesex  
TW11 8PA

Client's Order Number: P16265  
Works Order Number: 12511-02  
Date of Test: 7<sup>th</sup> to 17<sup>th</sup> May 2010

Attn.: Mr. Olivier Hus

**Specimens:** 1 off FT702LT Wind Sensor  
Serial No.: 2870-001  
Part No.: 22 (With Soldered Internal Connector)  
TRaC Stores No.: 24444  
Receipt Date: 19<sup>th</sup> April 2010

**Specification:** Steady State Humidity Test

Tested in accordance with BS EN 60068-2-78: 2002, Test Cab

Temperature: +40°C  
Relative Humidity: +93%  
Duration: 240 hours

**Procedure:** The specimen was attached to a support framework inside the test chamber and a aspirated hygrometer was positioned adjacent to the test specimen to record the ambient air conditions during the test. A representative of the customer then connected the specimen under test to function test equipment located external to the test chamber. The specimen was placed in an operational state by the customer and its operation was continuously logged by the customer's laptop. The specimen was tested in accordance with the specification and on completion of the test the specimen was visually inspected for any conspicuous signs of external damage or degradation.

TEST ENGINEER



D. Wheatley

Q.A. APPROVAL



D.K.Morris – Chief Test Engineer

Certified that the specimens detailed herein have been subjected to the tests as required by the order unless otherwise stated above. Our technical competence and quality control arrangements are in accordance with the conditions of our UKAS accreditation. No representation or warranty is given that the Tests performed under the terms of the Contract constitute, in themselves, a sufficient programme for the Customer's purpose, nor that the Customer's Equipment is suitable for any particular purpose. The contents of this Certificate shall not be reproduced, except in full, without the written approval of TRaC Global Limited.

**WARWICK**

Rothwell Road, Warwick, CV34 5JX, UK.

T +44 (0)1926 478478 F +44 (0)1926 478479 E test@tracglobal.com

www.tracglobal.com

Issue Date: 7<sup>th</sup> July 2010



0026



## *Certificate of Testing*

**Serial Number: 12511CC60A**

Page 2 of 2 Pages

**Results:**

The specimen completed the test with no conspicuous signs of external damage or degradation and the customer reported that the specimen had operated correctly throughout the test.